



中国认可
国际互认
校准
CALIBRATION
CNAS L0358

常州检验检测标准认证研究院

Changzhou Institute of Inspection Testing Standardization and Certification

校准证书

Calibration Certificate

证书编号: 721003233-011 号

Certificate No.



委托方

Customer

常州天策电子科技有限公司

委托方地址

Address of Customer

常州市武进区科教城创研港1号楼A座901室

样品名称

Name of Instrument

分布式薄膜压力传感器

型号/规格

Type /Specification

CN1911_500PSI

样品编号

Sample No.

17KROB50032

制造单位

Manufacturer

常州天策电子科技有限公司

(证书/报告专用章)

Stamp



校准

Calibrated by

李峥

核验

Checked by

凌年

批准

Approved by

李志峰

批准日期

Authorize Date

2021 - 10 - 15

样品接收日期

Date of Received

2021

年

10

月

15

日

Year

Month

Day

校准日期

Date of Calibrated

2021

年

10

月

15

日

Year

Month

Day

地址: 常州市武进区西湖路2-18号

Address

邮政编码: 213164

Post code

电话: 4008580800

Tel.

网址: <http://www.cztsi.com>

Website

监督电话: 0519-86290020

Tel. for Complain

传真: 0519-81002098

Fax.

校准结果

编号: 第 721003233-011 号
No.

Results of Calibration 第 3 页 共 3 页
Page of

校准点	理论力值	显示值	示值误差	示值允差	重复性	重复性允差	扩展不确定度 U_{rel}
(psi)	(N)	(psi)	(%)	(%)	(%)	(%)	(%)
100	16548	92.6	-7.4	±10.0	0.7	2.0	0.8
200	33096	186.4	-6.8		0.4		0.6
300	49644	311.1	3.7		1.2		1.0
400	66192	421.2	5.3		1.5		1.2
500	82740	541.0	8.2		0.9		0.8
备注: 试样受力面积为 24000 mm ²							

以下空白



注: 本校准结果只对所校准样品有效。

Note: The results are only responsible for the items calibration.

2. 未经本实验室书面批准, 不得部分复制本校准证书。

This certificate of calibration shall not be reproduced except in full, without the written approval of the laboratory.

3. 本证书未加盖证书/报告专用章无效。

This certificate will be invalid if it is issued without official stamp.